

CE Declaration of Conformity

**Manufacturer name
and address** Biotech AB
Råövägen 300
Onsala, Sweden

DEGASi® CLASSIC Product number: 0001-6XXX
MICRO
SEMI-PREP
GPC



TEST REPORT

Applicant: Biotech AB
Råövägen 300, 439 92 Onsala, Sweden

Description of sample: Biotech DEGASi®Prep, Biotech DEGASi®Prep+,
Biotech DEGASi®Compact, Biotech DEGASi®Compact open,
Biotech DEGASi®OEM, Biotech DEGASi®Classic,
Biotech DEGASi®GPC, Biotech DEGASi®Micro,
Biotech DEGASi®Semi-prep, Biotech DEGASi®High Flow


Date for test: 2015-09-02, 09-07, 09-24

Investigation requested: Test for compliance with EMC requirements of SS-EN 61326-1:2013.

Conclusion(s): The submitted products COMPLIED with the EMC requirements.
EMC testing was performed in accordance with the standard mentioned above and with relevant methods used by the test house.

We declare that the tested unit(s) is in accordance with the results and specifications in this report.
The test result relates only to the tested item(s).




Jan Linders
Authorized signatory
Licentiate of Engineering
2016-04-13
Date